

Vergleichsmessungen

Datum	Zeit	Temperatur (aussen)					Luftfeuchtigkeit (aussen)				
		Referenz					Referenz				
		Davis		Vaisala	Differenz	Korrektur	Davis	Vorschlag	Vaisala	Differenz	Korrektur
20.02.2007	08:24	-0.7		-0.48	0.22		91.0	100.0	100.00	9.00	
19.02.2007	21:15	4.4		4.05	0.35		83.0		95.60	12.60	
08.03.2007	07:45	5.8		5.76	0.04		85.0		94.41	9.41	
23.02.2007	22:50	5.7		5.55	0.15		79.0		91.13	12.13	
20.02.2007	19:17	5.5		4.99	0.51		80.0		90.80	10.80	
02.03.2007	20:15	5.6		5.57	0.03		78.0		86.03	8.03	
04.03.2007	23:23	4.7		4.81	-0.11		73.0		79.39	6.39	
04.03.2007	23:08	5.3		5.27	0.03		71.0		77.77	6.77	
08.03.2007	21:05	6.8		6.69	0.11		66.0		72.71	6.71	
03.03.2007	22:35	8.0		8.07	-0.07		65.0		68.85	3.85	
21.02.2007	23:40	8.3		8.34	-0.04		64.0		67.50	3.50	
04.03.2007	18:40	9.4		9.39	0.01		59.0		61.80	2.80	
03.03.2007	16:40	10.8		10.93	-0.13		57.0		58.52	1.52	
03.03.2007	17:20	10.4		10.62	-0.22		56.0		57.41	1.41	
04.03.2007	17:45	11.5		11.34	0.16		53.0	54.0	54.65	1.65	
10.03.2007	12:11	9.2		9.37	-0.17		52.0		52.10	0.10	
04.03.2007	17:00	13.7		13.68	0.02		48.0		46.60	-1.40	
04.03.2007	14:30	14.9		14.86	0.04		47.0		46.00	-1.00	
04.03.2007	16:28	14.6		14.54	0.06		46.0		44.72	-1.28	
11.03.2007	14:20	11.9		11.35	0.55		44.0		42.76	-1.24	

